


# VALIDATION OF STEAM STERILIZERS IN COMPLIANCE TO EN 285 AND ISO 17665

## ANNUAL PERFORMANCE REQUALIFICATION



EN 285, the European Large Steam Sterilizer standard [1], is the world's baseline authority for steam quality acceptance criteria. It is referenced in most national standards and in ISO 17665 [2]. With the release of EN 285:2013, the bar has been raised. The acceptance criteria are shown in the following table.

Steam Dryness	Non-condensable gases	Duperheat
>0.99 w/w*	<3.5% v/v	≤5K



# AS/NZS 4187:2014 Section 7: Validation

## 7.4 PERFORMANCE QUALIFICATION

- Stage of validation:
  - Exposes the products as specified to the sterilization process
  - Demonstrates equipment consistently operates in accordance with predetermined criteria
  - Yields a product that is sterile and meets specified requirements

## 7.5.2.(i)

- Performed using RMDs of each type of product families, if applicable heat penetration studies for each type of load.  
Further: Successful PQ = can be extrapolated to other loads that are covered by the reference load.

# AS/NZS 4187:2014 Section 7: Validation

## 7.4.5 STERILIZING PROCESSES

- Demonstrate the attainment of the required sterilizing conditions ON and THROUGHOUT a reusable medical device (RMD) within the specified sterilizer load.
  - Physical performance qualification
  - Microbial performance qualification
- For PQ of a moist heat sterilizing process the penetration time to all parts of an RMD shall be established and added to the sterilization holding time (refer to table 6.2).

# AS/NZS 4187:2014: Relevant Parts

**Table 6.2**

**TABLE 6.2  
COMMON HOLDING TIMES  
FOR SATURATED STEAM**

Temperature °C	Holding time min
121	15
126	10
132	4
134	3

NOTE: The correct correlation between temperature and pressure is necessary to ensure the presence of saturated steam. For pressure relationships, see Figure 6.1.

## **PQ for a moist heat must include:**

- Establishment of the heat penetration time using the most difficult to sterilize RMD or RMD Set. Penetration time must be added to the holding time of the defined temperature.
- Demonstration of microbial lethality of the process strategically positioned on most difficult to penetrate

# ISO 17665 - 3

ATTRIBUTE	CODE
DESIGN	a
WEIGHT	b
MATERIAL	c
STERILE BARRIER SYSTEM	d

STRUCTURE	CODE (A)	EXAMPLE
SOLID/HOLLOW	1	BOWLE/JUG/TRAY/CHISEL
PIN & BOX JOINTS	2	SCISSORS/FORCEPS
LUMEN	3	LAP SHEATH, SUCTIONS
POROUS	4	FILTERS
TUBING/MOVING PARTS TORTUOUS PATHS	5	POWER TOOL/HOSE/ENT DRILL/DENTAL HANDPIECE
LUMEN SURROUNDED - LARGE MASS	6	CANNULATED SCREW DRIVER, OBTURATOR, RACHET HANDLE
OTHER	7	FRE FILLED SYRINGE

MATERIAL	CODE (B)	EXAMPLE
METAL	1	SS/COPPER BASED ALLOYS/OTHER METAL COMBINATIONS
NON METAL	2	GLASS/PVC/PLYCARBONATE /SILICON

WEIGHT GRAMS	CODE (C)
LESS THAN 50	1
50 - 499	2
500 - 1999	3
2000 AND GREATER	4

SBS	CODE (D)
NONE	1
SINGLE WRAP/POUCH	2
DOUBLE WRAPPED IN POUCH/DOUBLE WRAPPED CONTAINER TRAY	3
COMBINATION OF 2 OR MORE SYSTEMS	4

# ISO 17665 – PRODUCT FAMILIES IDENTIFIED

PRODUCT FAMILIES					
PF	ATTRIBUTE				STEAM PENETRATION
	A	B	C	D	E
1					
29					

Equipment: 4 X Steam Sterilizers		
Product Family 1	134 C	5 Minutes
Product Family 2	134 C	10 Minutes
Product Family 3	134 C	18 Minutes
Product Family 4	121 C	20 Minutes

# AS/NZS 4187:2014 – PQ

Thermocouple placement within the sterilizer chamber is shown in the diagram below.




Figure 1.2 – Heat Distribution Profile:




Table 1: Heat Distribution Measurements:

Thermocouple Location	Temp (°C)	
1. Drain	135.28	
12. Front Lower Right	135.38	
9. Centre Lower Centre	135.42	
11. Front Centre Centre	135.49	
10. Front Upper Left	135.55	
5. Rear Lower Left	135.58	
3. Pack	135.59	
8. Centre Upper Centre	135.60	
6. Rear Centre Centre	135.68	
4. Above Pack	135.75	
7. Rear Upper Right	135.81	

# PRODUCT FAMILY

1




**PFNA Lock and Removal Tray I**



**GEN 2 Universal Patella Tray**



# Reference Load 1/Product Family 1




Load Description and Instrument Penetration times. (Also refer to diagram over page)

Part #	Description	Volume (L)	Penetration Time during sterilize phase			Location of Probe Number	Shape	Weight (kg)
			Cycle #16411	Cycle #16412	Cycle #16413			
1	Gen 2 Universal Penetration Tray	12*	<30s	<30s	<30s	Center Shelf, Front	Flask	6.4
2	Probe 2 Glass and Soap Tray					Center Shelf, Center Front	Chamber Heavy Duty	7.9
3	Probe 2 Cork and Reservoir	9	<30s	<30s	<30s	Lower Shelf, Center Rear	Chamber Heavy Duty	7.0
4	Microtome Holder					Lower Shelf, Rear		6.0
5	Humane Personnel	11*	<30s	<30s	<30s	Center Shelf, Front	Flask	6.4
6	Gen 2 Cutting Blocks					Center Shelf, Center Rear		6.4
7	I & M Spindle Attachments					Center Shelf, Center Rear		3.0
8	Gen 2 Femoral Trays 2 & 3 Left					Center Shelf, Rear		2.4
9	Gen 2 Tibial Tray 3-4					Center Shelf, Rear	Chamber Heavy Duty	2.6
10	I & M Spindle Attachments					Lower Shelf, Front		3.0
11	Gen 2 Tibial Tray 3-5 (single)					Lower Shelf, Center Front		3.0
12	Gen 2 Femoral Trays 3 & 2					Lower Shelf, Center		2.0
13	Gen 2 Tib Tray 2-8 (single)					Lower Shelf, Center Rear		2.0
14	Gen 2 Tib Tray 1-1 (single)					Lower Shelf, Rear		2.0
15	EA Vial Bottle Grouped Box	10*	<30s	<30s	<30s	Upper Shelf, Rear	Flask, Chamber	6.2


Total Weight - 54.8kg

# Penetration time graphical data – PF 1



# Reference Load 2/Product Family 2

Diagram 1.6



Prevac 2 Complex Instruments Load Layout – Sterilizer #1 – Cycle #16421-16423 Load Configuration  
Total Load Weight\*\* – 10.15kg

5.7 Complex Instrument Load Test Results

Program Parameters		Sterilizer Cycle	Test Cycle Sequence	Test Result
Program Name: 'Prevac 2'		#16421	1/3	PASS
Sterilization:	134°C for 15m	#16422	2/3	PASS
Drying:	50m dry	#16423	3/3	PASS

Load Description and Instrument Penetration times. (Also refer to diagram over page)

Pack #	Description	T/Coupe ID	Penetration Time during sterilis phase			Location of Pack in Chamber	Area	Weight (kg)
			Cycle #16421	Cycle #16422	Cycle #16423			
a	ASPS II 4mm Tray	3*	1m 52s	<30s	<30s	Lower Shelf, Front	Quadrat 75	2.0
b	ASPS II 6.5/8mm Tray	2*	<30s	<30s	<30s	Lower Shelf, Centre	Quadrat 75	2.4
c	Orion-Clage Cable System	11*	5m	6m 30s	4m 30s	Centre Shelf, Front	Inverted	6.0
d	Spacers Cumulated Screwdriver Instrument #114 (75)	10	<30s	4m 34s	3m 32s	Upper Shelf, Front	Quadrat Inverted	0.75
Total Weight: 10.15kg								

# PRODUCT FAMILY

## 2





**Osteo - Clage cable System**



**Synthes Cannulated Screw Driver**

# Penetration time graphical data – PF2




# PRODUCT FAMILY 3




**LARS ATVS 3.5**

**LARS ACL Tray**



# Reference load 3/Product family 3





## 5.5 Ultra Complex Instrument Load Test Results

Program Parameters		Sterilizer Cycle	Test Cycle Sequence	Test Result
Program Name: 'Prevac 2'		#16414	1/3	PASS
Sterilization:	134°C for 18m	#16415	2/3	PASS
Drying:	20m dry	#16416	3/3	PASS

Load Description and Instrument Penetration times. (Also refer to diagram over page).

Pack #	Description	T/Couple ID#	Penetration Time during sterilize phase			Location of Pack in Chamber	Verify	Weight (kg)
			Cycle #16414	Cycle #16415	Cycle #16416			
a	Lars ACL Tray	1*	10m 40s	8m 59s	9m 20s	Lower Shelf, Front	Intacture®	6.4
b	Lars ATVS 3.5 (Instrument provided by SCA for testing)	11*	12m 43s	10m 43s	11m 7s	Centre Shelf, Front	Chemistry Standard	0.5
Total Weight: 4.9kg								

# Penetration time graphical data – PF3





# PRODUCT FAMILY 4

## GEN 2 Universal Patella Tray




**Biotronik Pacemaker Lead**



# Reference load 4/Product family 4

Diagram 1.3



Prevac 3 Low Temp 121° Instruments Load Layout – (Sterilizer #1) – Cycle #16417 – 16420 Load Configuration  
Total Maximum Validated Load Weight – 29.45kg

BI Biological Indicator

PCD Process Challenge Device

5.8 Low Temp 121°C Load Test Results

Program Parameters		Sterilizer Cycle	Test Cycle Sequence	Test Result
Program Name: 'Prevac 3'		#16417	1/3	PASS
Sterilization:	121°C for 20m	#16419*	2/3	PASS
Drying:	80m dry	#16420	3/3	PASS

\*Cycle 16418 – Upper rack family, test sequence not required.

Load Description and Instrument Penetration Times. (Also refer to diagram over page)

Pack #	Description	V-Curve ID	Penetration Time during cycle phase			Location of Pack in Chamber	Size	Weight (kg)
			Cycle #16417	Cycle #16418	Cycle #16420			
a	Gen 2 Ultrasonic	11*	<30s	<30s	<30s	Lower Shelf, Front	5.4	
b	FFKA 8 Open and Insert					Lower Shelf, Centre	7.0	
c	FFKA 8 Lock and Removal					Lower Shelf, Rear	7.0	
d	Gen 2 Cutting Blocks 3 R & DFC	11*	<30s	<30s	<30s	Centre Shelf, Front	4.4	
e	Mixed Instruments	10*	<30s	<30s	<30s	Upper Shelf, Front	4.5	
						Upper Shelf, Rear	4.0	

Total Weight = 29.45kg



# IN CONCLUSION

Compliance to standards is necessary to maintain product realization. In this case, RMDs must be reprocessed according to its specific requirements of exposure to predetermined values necessary to attain lethality and meet minimum sterility assurance level.

